

Infrared thermometer



Société BIZLINE SAS
13 boulevard du Fort de Vaux
75017 Paris
FRANCE

 WE, BIZLINE, ASSUME FULL RESPONSIBILITY FOR COMPLIANCE OF THE FULL REFERENCE BIZ 790 395 WITH THE PROVISIONS OF THE EC DIRECTIVE(S) GIVEN BELOW :

- LVD/Low voltage directive 2014/35/EU
- EMC 2014/30/EU
- RoHS 2011/65/EU

 WE DECLARE COMPLIANCE OF THE PRODUCT WITH THE HARMONISED STANDARDS GIVEN BELOW :

- LVD : EN 60825-1:2014
- EMC : EN 61326-1 :2013 EN 61326-2-2 :2013

M. Nicolas MOUGEL
Technical, Quality & Laboratory Manager

Paris, 30/06/2024

A handwritten signature in black ink, appearing to be 'M. Nicolas MOUGEL'.